

Accurc

Immersion and external disinfection guidelines.



The following are approved for use with RIVANNA Accuro systems:

| ULTRASOUND GELS | CLEANERS/DISINFECTANTS | HIGH-LEVEL DISINFECTANTS* |
|-----------------|------------------------|---------------------------|
| Aquagel | Cavi Wipes | Cidex* |
| Aquasonic 100 | Transeptic Spray | |
| Aquasonic Clear | Sono wipes | |
| Clear Image | P.I. Spray II | |
| Clear Wave | Protex Wipes | |
| Scan | Sani-cloth HB Wipes | |
| Ultraphonic | Sani-cloth Plus Wipes | |

*Not to be used for immersion disinfection, external disinfection only.

Rivanna Medical, LLC Charlottesville, Virginia USA

T: 800-645-7508

© Copyright, Rivanna Medical, LLC, all rights reserved. Unauthorized copying of this document, in addition to infringing copyright, might reduce the ability of RIVANNA to provide accurate and current information to users.